


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-5B-C-M-P*
S.N.: *3892*
Date: *SEPT 287*



- a) Output signal amplitude:
0 TO +48 AMPS
- b) Pulse width:
25 TO 500 NSER
- c) Rise time:
≤ 10 NSER
- d) Fall time:
≤ 10 NSER
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz
40

R_L = 0.7 Ω
50 dB ATTEN
≈ 32 VOLTS/DIV
≈ 46 AMPS/DIV