


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO 1-P*

S.N.: *3899*

Date: *SEPT 3 1987*



- a) Output signal amplitude: *+25 TO 100. VPPS*
- b) Pulse width: *20 ± 5 nSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1 KHz.*
- f) Jitter, stability: *OK*
- g) Prime power:

+15 ± 1.0 VOLTS

250 mA (MAX)
AJ