



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-7A - C - N*  
S.N.: *3901*  
Date: *SEPT 9 87*



Ⓐ  $R_L = 0.5 \Omega$   
 $100 \text{ KHz}$

- a) Output signal amplitude:  
*0 TO -40 AMPS*
- b) Pulse width:  
*0.2 TO 200  $\mu$ SEC.*
- c) Rise time:  
 *$\leq 100 \text{ NS}$*
- d) Fall time:  
 *$\leq 100 \text{ NS}$*
- e) PRF:  
*MAX PW: 0 TO 50 Hz.*  
*PW  $\leq 10 \text{ nsec}$ : 0 TO 1 KHz.*
- f) Jitter, stability:  
*OK*



Ⓑ  $R_L = 0.5 \Omega$   
 $50 \text{ Hz}$

- g) Prime power:  
*120/240 V*  
*50-60 Hz*
- [Signature]*