


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2-PS-P  
S.N.: 3904  
Date: OCT 6 87



70 db ATTEN  
≈ 320 VOLTS/DIV  
2.0 KHz.  
( $R_L = 50 \Omega$ )

- a) Output signal amplitude:  
0 TO +350 V
- b) Pulse width:  
5 TO 100 NSEC
- c) Rise time:  
 $\leq 2$  NSEC
- d) Fall time:  
 $\leq 3$  NSEC
- e) PRF:  
0 TO 5 KHz.
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50-60 Hz.

*[Handwritten signature]*