


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-MN-1403-05-C-P*

S.N.: *3905*

Date: *SEPT 18 87*



*50 db ATTEN
≈ 32 VOLTS/DIV
100 KHz*

- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
0.4 TO 2.0 NSEC
- c) Rise time:
150 PSEC
- d) Fall time:
≤ 250 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120 / 240V
50-60 Hz*

[Signature]