


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-MPT-P*

S.N.: *3912*

Date: *SEPT 16 87*



- a) Output signal amplitude:  
*> +30 VOLTS*
- b) Pulse width:  
*2 / NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 dB ATTEN  
32 VOLTS/DIV  
500 KHz*

*+15 ± 1.0 VOLTS  
125 mA (MAX)*