


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-MPT-P.

S.N.: 3913

Date: SEPT. 16 87



- a) Output signal amplitude:  
7, +30 VOLTS
- b) Pulse width:  
≤ 1 NSEC
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1.0 MHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 ± 1.0 VOLT  
125 mA (MAX)

50 db ATTEN  
32 VOLTS/DIV  
500 KHZ