


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-MPT-P

S.N.: 3915

Date: SEPT 16 87



a) Output signal amplitude:

$\approx +30$ VOLTS

b) Pulse width:

≤ 1 NSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1.0 MHZ

f) Jitter, stability:

OK

g) Prime power:

$+15 \pm 1.0$ VOLT

125 mA (MAX)

50 db ATTEN
32 VOLTS/DIV
500 KHZ