


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-MPT-P.

S.N.: 3917

Date: SEPT. 16 87



- a) Output signal amplitude:
 $\approx +30$ VOLTS.
- b) Pulse width:
 ≤ 1 NSEC
- c) Rise time:
 NA
- d) Fall time:
 NA
- e) PRF:
 0 TO 1.0 MHZ
- f) Jitter, stability:
 OK
- g) Prime power:

$+15 \pm 1.0$ VOLT

125 mA (MAX)

50 db ATTEN
32 VOLTS/DIV
500 KHZ