


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-MPT-P.

S.N.: 3919

Date: SEPT. 16 87



50 db ATTEN  
32 VOLTS/DIV  
500 KHZ

- a) Output signal amplitude:  
     $\geq +30$  VOLTS.
- b) Pulse width:  
     $\leq 1$  NSEC
- c) Rise time:  
    NA
- d) Fall time:  
    NA
- e) PRF:  
    0 TO 1.0 MHZ
- f) Jitter, stability:  
    OK
- g) Prime power:  
     $+15 \pm 1.0$  VOLT  
    125 mA (MAX)