


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-MPT-R

S.N.: 3920

Date: SEPT 16 87



50 db ATTEN
32 VOLTS/DIV
500 KHZ

- a) Output signal amplitude:
7/ +30 VOLTS.
- b) Pulse width:
≤ 1 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHZ
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLT
125 mA (MAX)