


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-MPT-P.

S.N.: 3921

Date: SEPT 16 87



50 db ATTEN
32 VOLTS/DIV
500 KHz

- a) Output signal amplitude:
 $\geq +30$ VOLTS
- b) Pulse width:
 ≤ 1 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
 $+15 \pm 1.0$ VOLT
125 mA (MAX)