


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5-1-C-P

S.N.: 3926

Date: SEPT 28 87



40 db ATTEN
10 VOLTS/DIV
1.0 nsec

- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
≤ 130 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60 Hz

[Signature]