


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVMN-1-N .

S.N.: 3931


Date: OCT 7 '87



- a) Output signal amplitude:
0 TO -5V
- b) Pulse width:
0.3 TO 5 NSEC
- c) Rise time:
 ≤ 100 PSEC
- d) Fall time:
 ≤ 200 PSEC
- e) PRF:
0 TO 50 MHZ
- f) Jitter, stability:
OK
- g) Prime power:
 $+24 \pm 1.0V$
500 mA (MAX)

(A)

40 db ATTEN
 $\therefore 5$ VOLTS/DIV
 ≈ 50 MHZ



(B)

AS (A) BUT
1 NSEC/DIV