


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2A-P
S.N.: 3924
Date: OCT 5 1987



60 db ATTEN
100 VOLTS / DIV
5.0 KHz
 $R_L = 50 \Omega$

- a) Output signal amplitude:
0 TO +170 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
 ≤ 2 NS
- d) Fall time:
 ≤ 2 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 \pm 1.0 VOLTS
400 mA (MAX)