


PULSE GENERATOR

PERFORMANCE CHECK

Model: APR-1-PW-C-P

S.N.: 3968

Date: OCT 29 87



$R_c = 50 \Omega$
PRF = 1 kHz

- a) Output signal amplitude:
0 TO +200V
- b) Pulse width:
0.1 TO 5.0 μ S
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF:
0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz