


PULSE GENERATOR
PERFORMANCE CHECK


Model: *ANMN-1-C-P-OT*

S.N.: *3976*

Date: *NOV 4 1987*



① 40 db ATTEN
∴ 5.0 VOLTS/DIV
12.0 MHz



② AS ① BUT
≈ 50 MHz

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.3 TO 5.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 MHz