


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ATM-1-D-EN-D*

S.N.: *3980*

Date: *SEPT 25 87*



*40 db ATTEN*  
*5 VOLTS/DIV*  
*5.0 MHz*

- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*0.2 TO 6.0 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC.*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+24 ± 1.0 VOLTS*  
*500 mA (MAX)*

*[Signature]*