


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVM-1-P-EW-D

S.N.: 3981

Date: SEPT 25 '87



40 db ATTEN
∴ 5 VOLTS/DIV
5.0 MHZ

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHZ
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0 VOLTS
500 mA (MAX)