


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVM-1-P-EW-D

S.N.: 3982

Date: SEPT 25 '87



40 db ATTEN  
∴ 5 VOLTS/DIV  
5.0 MHz

- a) Output signal amplitude:  
0 TO +5 VOLTS
- b) Pulse width:  
0.2 TO 6.0 NSEC
- c) Rise time:  
≤ 100 PSEC
- d) Fall time:  
≤ 135 PSEC
- e) PRF:  
0 TO 25 MHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
+24 ± 1.0 VOLTS  
500 mA (MAX)