


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVM-1-P-EW-D

S.N.: 3983

Date: SEPT 25 '87



- a) Output signal amplitude:  
0 TO +5 VOLTS
- b) Pulse width:  
0.2 TO 6.0 NSEC
- c) Rise time:  
 $\leq 100$  PSEC
- d) Fall time:  
 $\leq 135$  PSEC
- e) PRF:  
0 TO 25 MHZ
- f) Jitter, stability:

OK

- g) Prime power:  
+24  $\pm$  1.0 VOLTS  
500 mA (MAX)

40 db ATTEN  
 $\therefore$  5 VOLTS/DIV  
5.0 MHZ

