


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APM-2-CA-EN-P*

S.N.: *4001*

Date: *NOV 26 87*



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NS
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS / DIV
7.0 MHz

+24 ± 1.0 VOLTS
500 mA (MAX)
[Signature]