


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-TN-1

S.N.: 4004

Date: OCT. 16/87



- a) Output signal amplitude:
- 230 VOLTS
WITH +40 V OFFSET
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time: ≤ 2 NSEC
- d) Fall time: ≤ 3 NSEC
- e) PRF: 0 TO 5 KHZ
- f) Jitter, stability: OK
- g) Prime power: 100/240 V
50-60 Hz.

60 db ATTEN
100 VOLTS / DIV.
2.0 KHZ