


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-TN-1.

S.N.: 4005

Date: OCT. 16 /87



- a) Output signal amplitude:
-230 VOLTS
- b) Pulse width:
WITH +40 V OFFSET
5 TO 100 NSEC
- c) Rise time: ≤ 2 NSEC
- d) Fall time: ≤ 3 NSEC
- e) PRF: 0 TO 5 KHZ
- f) Jitter, stability: OK
- g) Prime power: 100/240 V
50-60 Hz.

60 db ATTEN

100 VOLTS / DIV.

2.0 KHZ