


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MVM-3-P*
S.N.: *4009*
Date: *NOV 26 87*



- a) Output signal amplitude:
0 TO +15 VOLTS.
- b) Pulse width:
2 TO 15 NSEC
- c) Rise time:
< 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS/DIV
7.0 MHz

424 ± 1.0 VOLT
500 mA (MAX)