


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ARM -3 -P*
S.N.: *4010*
Date: *NOV 26 87*



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
2 TO 15 NS
- c) Rise time:
≤ 150 PS
- d) Fall time:
≤ 600 PS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS DIV
7.0 MHz.

+24 ± 1.0 VOLTS
500 mA (MAX)