


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AMK-B3-W-PN-C-R3*

S.N.: *4011*

Date: *NOV 19 87*



Ⓐ  $P_{out} R_L = 50\Omega$   
*5 KHz*




Ⓑ  $P_{out}$  60 dB ATTEN  
100 VOLTS/DIV  
(RISE TIME)



Ⓒ  $N_{out} R_L = 50\Omega$   
*5 KHz*

- a) Output signal amplitude:  
*0 TO ± 100 VOLTS*
- b) Pulse width:  
*0.1 TO 5 μSEC*
- c) Rise time:  
*≤ 5 NS*
- d) Fall time:  
*≤ 5 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*



Ⓓ  $N_{out}$  60 dB ATTEN  
100 VOLTS/DIV  
(RISE TIME)