


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AP-M-1-C-PN-M*

S.N.: *4016*

Date: *OCT 1 87*



*40 db ATTEN*  
*10 VOLTS/DIV*  
*1.0 MHz*

- a) Output signal amplitude:  
*0 TO ±10 VOLTS*
- b) Pulse width:  
*0.2 TO 4.0 nSEC.*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 135 pSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50 - 60 Hz*

*[Signature]*