


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-MV-1-T*

S.N.: *4024*

Date: *NOV 23 87*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS DIV
1.0 MHz

+15 ± 1.0 VOLTS
250 mA (MAX)