



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-3-C-P-PN-OT-M*
S.N.: *4027*
Date: *DEC 11 87*



① *40 db ATTEN*
10 VOLTS/DIV
1.0 MHz



② *AS ① BUT*
1 nsec/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO ± 20 VOLTS
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
≤ 200 PSEC
- d) Fall time:
≤ 300 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz
[Signature]