


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-MPT-P

S.N.: 4036

Date: DEC 2, 1987



- a) Output signal amplitude:
 ≥ 30 VOLTS
- b) Pulse width:
 ≤ 1.0 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
 $+15 \pm 1.0$ VOLT
150 mA (MAX)

50 db ATTEN
 $\therefore 32$ VOLTS/DIV
500 KHZ