


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-MPT-P

S.N.: 4037

Date: DEC 2, 1987



a) Output signal amplitude:

≈ 30 VOLTS

b) Pulse width:

≤ 1.0 NSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

$\pm 15 \pm 1.0$ VOLT

150 mA (MAX)

50 db ATTEN

$\therefore 32$ VOLTS/DIV

500 KHz