


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MM-1-C-EN-P*

S.N.: *4038*

Date: *DEC 21 87*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 KHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
5 VOLTS/DIV
≈ 12 MHz

120/240V
50-60 Hz