


PULSE GENERATOR

PERFORMANCE CHECK

Model: PVL-TN-1

S.N.: 4040

Date: OCT 16 87



- a) Output signal amplitude:
-230 VOLTS
- b) Pulse width: WITH 440 V OFFSET
5 TO 100 NSEC.
- c) Rise time:
≤ 2 NSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

60 dB ATTENS

100 VOLTS/DIV

2.0 KHz

120/240 V

50-60 Hz