


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-2-C-D-M-N-PN*

S.N.: *4043*

Date: *DEC 4 87*



- a) Output signal amplitude:
0 TO \pm 15 VOLTS
- b) Pulse width:
0.2 TO 2.0 μ SEC
- c) Rise time:
 \leq 100 PSEC
- d) Fall time:
 \leq 130 PSEC
- e) PRF:
0 TO 25 kHz
- f) Jitter, stability:
OK
- g) Prime power:

*40 db ATTEN
10 VOLTS/DIV
 \approx 10 MHz*

*120/240V
50-60 Hz*
