


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANMP-2-PS-ETA EW -P-PN*

S.N.: *4063*

Date: *DEC 15 1980*



*40 dB ATTEN
10 VOLT/DIV
1.0 MHz*

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*

[Signature]