



PULSE GENERATOR
PERFORMANCE CHECK

Model: *ARR-3 PW-PS-PN*
S.N.: *4067*
Date: *DEC 4 1987*



- a) Output signal amplitude:
0 to ± 200 VOLTS
- b) Pulse width:
0.1 to 100 µSEC
- c) Rise time:
≤ 10 nSEC
- d) Fall time:
≤ 10 nSEC
- e) PRF:
MIN PW: 1 kHz
MAX PW: 50 Hz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

Ⓐ *P_{out}*
R_L = 50 Ω
50 Hz PRF



Ⓑ *N_{out}*
R_L = 50 Ω
500 Hz PRF

[Signature]