


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APP-1-C-P*

S.N.: *4068*

Date: *DEC 8 1987*



(A) A out
40 db ATTEN
10 VOLTS/DIV
1.0 MHz



(B) B out
40 db ATTEN
10 VOLTS/DIV
1.0 MHz

- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 100 NSER
- c) Rise time: (Two Ranges)
6 100 PSEZ
- d) Fall time:
6 135 PSEZ
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50 - 60 Hz