


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APM-1-C-P*

S.N.: *4073*

Date: *JAN 6 1988*



- a) Output signal amplitude:
0 TO +5V
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V

50-60 Hz

[Signature]

40 dB ATTEN

5.0 VOLT/DIV

≈ 12 MHz