


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-M-1-C-PN-05-M*

S.N.: *4074*

Date: *DEC 7 87*



*40 dB ATTEN
10 VOLTS/DIV
1.0 MHz*

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*