


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PAUL-2-PS-T*

S.N.: *4076*

Date: *JAN 18 88*



- a) Output signal amplitude:
0 TO + 350V
- b) Pulse width:
5 TO 100 NSER
- c) Rise time:
≤ 2 NSER
- d) Fall time:
≤ 3 NSER
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

70 db ATTEN
≈ 320 VOLTS PIV
2 KHZ.

120/240V
50-60 HZ
[Signature]