


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-MPT - P .

S.N.: 4098

Date: DEC 3 1987



50 db ATTEN  
∴ 32 VOLTS/DIV  
500 KHZ

- a) Output signal amplitude:  
 $\geq 30 \text{ VOLTS}$
- b) Pulse width:  
 $\leq 1 \text{ NSEC}$
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
 $+15 \pm 1.0 \text{ VOLTS}$   
150 mA (MAX)