

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVL-TN-1

S.N.:

4099

Date:

NOV 10 87



- a) Output signal amplitude:  
-230 VOLTS  
WITH +40 DC OFFSET
- b) Pulse width:  
5 TO 100 NSER.
- c) Rise time:  
≤ 2 NSER
- d) Fall time:  
≤ 3 NSER
- e) PRF:  
0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:

60 db ATTEN  
100 VOLTS/DIV  
2.0 KHz

120/240 V  
50-60 Hz