


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-1-PW-HA1*


S.N.: *4103*

Date: *DEC 16 87*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width:
TO 20 ns
0.1 TO 5.0 USEC
- c) Rise time:
< 20 NSEC
- d) Fall time:
< 20 NSEC
- e) PRF:
0 TO 200 Hz
- f) Jitter, stability:
OK
- g) Prime power:

① *R_i = 20 Ω*
200 Hz
I ≈ 2.5 AMP/DIV



+24 ± 1.0 VOLTS
500 mA (MAX)

[Signature]

③ AS ① BUT
100 NSEC/DIV