



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1C-P*  
S.N.: *4124*  
Date: *MAR 15 88*



① *40dB ATTEN*  
*50V/DIV*  
*= 25 MHz*



② *AS @ OUT*  
*5 nS/DIV*

- a) Output signal amplitude:  
*0 TO 15 VOLTS*
- b) Pulse width:  
*0.2 TO 6.0 nS*
- c) Rise time:  
*≤ 10 pS*
- d) Fall time:  
*≤ 135 pS*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*