



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMM-3-C-M-PN*
S.N.: *4129*
Date: *FEB 12 1988*



*(A) 40 dB ATTEN
-10 VOLT/DIV
≈ 5.0 MHz*



*(B) AS (A) BUT
≈ 25 MHz*

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
1 TO 5 NSEC
- c) Rise time:
≤ 300 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*120/240V
50-60 Hz*
[Signature]