


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ML-TN-1*

S.N.: *4140*

Date: *JUNE 24 '88*



- a) Output signal amplitude:  
*-230 VOLTS*
- b) Pulse width:  
*5 TO 100 NSOR*
- c) Rise time:  
*6 2 NSOR*
- d) Fall time:  
*5 3 NSOR*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240V*

*60 db ATTEN*

*100 VOLTS/DIV*

*200 KHZ*

*50 - 60 KHZ*

*A*