


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVC-TN-1

S.N.: 4141

Date: JUNE 24 88



- e) Output signal amplitude:
-230 VOLTS
WITH +40 V OFFSET
- f) Pulse width:
5 TO 100 NS
- g) Rise time:
≤ 2 NS
- h) Fall time:
≤ 3 NS
- i) PRF:
0 TO 5 KHZ
- j) Jitter, stability:
- k) Prime power:
OK
120/240 V
50-60 Hz

60 dB ATTEN
100 VOLTS/DIV
2.0 KHZ

[Signature]