


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-AN-2-P*


S.N.: *4142*

Date: *FEB 17 1988*



- a) Output signal amplitude:  
*0 TO +10 VOLTS.*
- b) Pulse width:  
*5 TO 50 NS*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 50 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 db ATTEN  
10 VOLTS/DIV  
50 KHZ*

*+15 ± 1.0 VOLT.  
300 mA (MAX)  
*