


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVRLE-ITT9A-05*

S.N.: *4159*

Date: *FEB 15 1988*



a) Output signal amplitude:
-200 V TO \approx 1K.

b) Pulse width:
20 TO 100 NS

c) Rise time:
 \leq 5 NS

d) Fall time:
 \leq 5 NS

e) PRF:
0 TO 1 KHz

f) Jitter, stability:
OK

g) Prime power:
*120/240V
50-60 Hz*

*$R_L = 10K.$
PRF = 1.0 KHz.*

h) OFFSET:
*0 TO \pm 50
VOLTS.*